

Arial (M901) 70 um	Total mL/h	Arial (M901) 100 um	Total mL/h
flow rate of 1.0	0.56	flow rate of 1.0	1.12
flow rate of 2.0	1.06	flow rate of 2.0	1.53
flow rate of 3.0	1.50	flow rate of 3.0	2.01
flow rate of 4.0	2.02	flow rate of 4.0	2.38
flow rate of 5.0	2.54	flow rate of 5.0	2.83
flow rate of 6.0	2.96	flow rate of 6.0	3.26
flow rate of 7.0	3.47	flow rate of 7.0	3.62
flow rate of 8.0	3.95	flow rate of 8.0	3.97
<u>flow rate of 9.0</u>	<u>4.37</u>	<u>flow rate of 9.0</u>	<u>4.41</u>
flow rate of 10.0	4.82	flow rate of 10.0	4.70
flow rate of 11.0	5.24	flow rate of 11.0	5.10
Arial (M901) 85 um	Total mL/h	Arial (M901) 130 um	Total mL/h
flow rate of 1.0	1.06	flow rate of 1.0	1.31
flow rate of 2.0	1.48	flow rate of 2.0	1.78
flow rate of 3.0	1.95	flow rate of 3.0	2.21
flow rate of 4.0	2.65	flow rate of 4.0	2.48
flow rate of 5.0	2.87	flow rate of 5.0	3.00
flow rate of 6.0	3.35	flow rate of 6.0	3.44
flow rate of 7.0	3.68	flow rate of 7.0	3.81
flow rate of 8.0	4.06	flow rate of 8.0	4.23
<u>flow rate of 9.0</u>	<u>4.49</u>	<u>flow rate of 9.0</u>	<u>4.64</u>
flow rate of 10.0	4.99	flow rate of 10.0	4.96
flow rate of 11.0	5.34	flow rate of 11.0	5.36

Ariall (T105) 70 um	Total mL/h	Ariall (T105) 100 um	Total mL/h
flow rate of 1.0	0.76	flow rate of 1.0	1.01
flow rate of 2.0	1.25	flow rate of 2.0	1.14
flow rate of 3.0	1.66	flow rate of 3.0	1.59
flow rate of 4.0	2.03	flow rate of 4.0	2.06
flow rate of 5.0	2.58	flow rate of 5.0	2.52
flow rate of 6.0	2.89	flow rate of 6.0	2.98
flow rate of 7.0	3.29	flow rate of 7.0	3.33
flow rate of 8.0	3.86	flow rate of 8.0	3.83
<u>flow rate of 9.0</u>	<u>4.22</u>	<u>flow rate of 9.0</u>	<u>4.28</u>
flow rate of 10.0	4.95	flow rate of 10.0	4.70
flow rate of 11.0	5.60	flow rate of 11.0	4.78
Ariall (T105) 85 um	Total mL/h	Ariall (T105) 130 um	Total mL/h
flow rate of 1.0	0.96	flow rate of 1.0	0.91
flow rate of 2.0	1.49	flow rate of 2.0	1.30
flow rate of 3.0	1.93	flow rate of 3.0	1.75
flow rate of 4.0	2.08	flow rate of 4.0	2.12
flow rate of 5.0	2.51	flow rate of 5.0	2.70
flow rate of 6.0	3.02	flow rate of 6.0	2.99
flow rate of 7.0	3.47	flow rate of 7.0	3.41
flow rate of 8.0	3.98	flow rate of 8.0	3.99
<u>flow rate of 9.0</u>	<u>4.46</u>	<u>flow rate of 9.0</u>	<u>4.48</u>
flow rate of 10.0	4.87	flow rate of 10.0	4.91
flow rate of 11.0	5.23	flow rate of 11.0	5.72

CAGT ArialIII (M903) 70 um	Total mL/h	CAGT ArialIII (M903) 130 um	Total mL/h
flow rate of 1.0	0.56	flow rate of 1.0	0.79
flow rate of 2.0	0.92	flow rate of 2.0	1.13
flow rate of 3.0	1.26	flow rate of 3.0	1.61
flow rate of 4.0	1.67	flow rate of 4.0	2.02
flow rate of 5.0	2.10	flow rate of 5.0	2.35
flow rate of 6.0	2.47	flow rate of 6.0	2.76
flow rate of 7.0	2.89	flow rate of 7.0	3.23
flow rate of 8.0	3.33	flow rate of 8.0	3.71
<u>flow rate of 9.0</u>	<u>3.79</u>	<u>flow rate of 9.0</u>	<u>4.09</u>
flow rate of 10.0	4.10	flow rate of 10.0	4.49
flow rate of 11.0	4.47	flow rate of 11.0	4.80

CAGT ArialIII (M903) 100 um	Total mL/h
flow rate of 1.0	0.62
flow rate of 2.0	0.98
flow rate of 3.0	1.38
flow rate of 4.0	1.76
flow rate of 5.0	2.18
flow rate of 6.0	2.63
flow rate of 7.0	3.01
flow rate of 8.0	3.43
<u>flow rate of 9.0</u>	<u>3.85</u>
flow rate of 10.0	4.27
flow rate of 11.0	4.52