

70um	85um	100um	130um

Aria I (M901) 70um	Total mL/h	Aria I (M901) 85um	Total mL/h
flow rate of 1.0	1.13	flow rate of 1.0	1.68
flow rate of 2.0	1.44	flow rate of 2.0	2.06
flow rate of 3.0	2.02	flow rate of 3.0	2.39
flow rate of 4.0	2.52	flow rate of 4.0	2.93
flow rate of 5.0	2.93	flow rate of 5.0	3.37
flow rate of 6.0	3.48	flow rate of 6.0	3.79
flow rate of 7.0	3.70	flow rate of 7.0	4.22
flow rate of 8.0	4.39	flow rate of 8.0	4.64
<u>flow rate of 9.0</u>	<u>4.80</u>	<u>flow rate of 9.0</u>	<u>5.26</u>
flow rate of 10.0	5.30	flow rate of 10.0	5.68
flow rate of 11.0	5.84	flow rate of 11.0	6.24
Aria I (M901) 100um	Total mL/h	Aria I (M901) 130um	Total mL/h
flow rate of 1.0	1.78	flow rate of 1.0	2.50
flow rate of 2.0	1.90	flow rate of 2.0	3.06
flow rate of 3.0	2.64	flow rate of 3.0	3.58
flow rate of 4.0	2.90	flow rate of 4.0	3.96
flow rate of 5.0	3.79	flow rate of 5.0	4.32
flow rate of 6.0	3.96	flow rate of 6.0	4.92
flow rate of 7.0	4.56	flow rate of 7.0	5.04
flow rate of 8.0	4.82	flow rate of 8.0	5.44
<u>flow rate of 9.0</u>	<u>5.16</u>	<u>flow rate of 9.0</u>	<u>5.82</u>
flow rate of 10.0	5.28	flow rate of 10.0	6.42
flow rate of 11.0	5.52	flow rate of 11.0	6.83

70um	85um	100um	130um

Aria II (T105) 70 um	Total mL/h	Aria II (T105) 85 um	Total mL/h
flow rate of 1.0	0.52	flow rate of 1.0	1.04
flow rate of 2.0	0.97	flow rate of 2.0	1.49
flow rate of 3.0	1.46	flow rate of 3.0	1.93
flow rate of 4.0	1.90	flow rate of 4.0	2.52
flow rate of 5.0	2.70	flow rate of 5.0	2.90
flow rate of 6.0	3.29	flow rate of 6.0	3.60
flow rate of 7.0	3.70	flow rate of 7.0	3.83
flow rate of 8.0	3.82	flow rate of 8.0	4.50
<u>flow rate of 9.0</u>	<u>3.98</u>	<u>flow rate of 9.0</u>	<u>4.92</u>
flow rate of 10.0	4.20	flow rate of 10.0	5.36
flow rate of 11.0	4.42	flow rate of 11.0	5.89
Aria II (T105) 100 um	Total mL/h	Aria II (T105) 130 um	Total mL/h
flow rate of 1.0	1.61	flow rate of 1.0	2.27
flow rate of 2.0	1.68	flow rate of 2.0	2.92
flow rate of 3.0	2.21	flow rate of 3.0	3.12
flow rate of 4.0	2.64	flow rate of 4.0	3.41
flow rate of 5.0	2.98	flow rate of 5.0	3.96
flow rate of 6.0	3.46	flow rate of 6.0	4.37
flow rate of 7.0	3.84	flow rate of 7.0	4.92
flow rate of 8.0	4.18	flow rate of 8.0	5.03
<u>flow rate of 9.0</u>	<u>4.49</u>	<u>flow rate of 9.0</u>	<u>5.71</u>
flow rate of 10.0	4.92	flow rate of 10.0	5.95
flow rate of 11.0	5.26	flow rate of 11.0	6.67

70um	85um	100um	130um

CAGT ArialIII (M903) 70 um	Total mL/h	CAGT ArialIII (M903) 130 um	Total mL/h
flow rate of 1.0	1.42	flow rate of 1.0	1.45
flow rate of 2.0	1.69	flow rate of 2.0	1.88
flow rate of 3.0	2.23	flow rate of 3.0	2.36
flow rate of 4.0	2.46	flow rate of 4.0	3.08
flow rate of 5.0	2.93	flow rate of 5.0	3.36
flow rate of 6.0	3.38	flow rate of 6.0	3.71
flow rate of 7.0	3.71	flow rate of 7.0	4.22
flow rate of 8.0	4.13	flow rate of 8.0	4.74
<u>flow rate of 9.0</u>	<u>4.49</u>	<u>flow rate of 9.0</u>	<u>5.16</u>
flow rate of 10.0	5.20	flow rate of 10.0	5.54
flow rate of 11.0	5.39	flow rate of 11.0	5.86

CAGT ArialIII (M903) 100 um	Total mL/h
flow rate of 1.0	1.44
flow rate of 2.0	1.76
flow rate of 3.0	2.18
flow rate of 4.0	2.69
flow rate of 5.0	3.11
flow rate of 6.0	3.49
flow rate of 7.0	3.72
flow rate of 8.0	4.37
<u>flow rate of 9.0</u>	<u>4.78</u>
flow rate of 10.0	5.15
flow rate of 11.0	5.74